



8 May 2024

To: Recipients of CLSI EP15-Ed3-IG1, 1st ed. and CLSI EP15-Ed3-IG2, 1st ed.  
From: Jennifer K. Adams, MLS(ASCP), MSHA  
Vice President, Standards and Quality  
Subject: Correction

This notice is intended to inform users of a correction made to CLSI EP15-Ed3-IG1, *User Verification of Precision Implementation Guide*, 1st ed. and CLSI EP15-Ed3-IG2, *User Verification of Bias (Trueness) Implementation Guide*, 1st ed. The correction is described below.

#### **CLSI EP15-Ed3-WB**

In the “Data Entry” tab of the worksheet (CLSI EP15-Ed3-WB), the standard errors of the mean (SE(mean)) for samples 2 (cell J39) and 3 (cell J52) were erroneously calculated from the within-lab and within-run imprecision estimates expressed as % CV. The errors in the calculations were corrected so the SE(mean) is correctly calculated from the within-lab and within-run imprecision estimates expressed in concentration units.

If you require any additional clarification regarding these corrections, please contact CLSI Customer Service ([customerservice@clsi.org](mailto:customerservice@clsi.org)).